



Based on the Assistance Measure to Strengthen the Capacities of the Balkan Medical Task Force (BMTF) financed by the European Peace Facility through the ITF Enhancing Human Security (ITF), ITF would like to announce the following:

### **INVITATION NOTICE FOR SUBMISSION OF OFFERS FOR THE DENTAL MODULE**

**1. Publication reference:**

Provision of the Dental Module

**2. Procedure:**

Participation is open on equal terms to private business entities (companies)

**3. Contracting Authority:**

ITF Enhancing Human Security (ITF)

**4. Description of the Procurement of Equipment:**

Please find detailed specification of the *Provision of the Dental Module* in Annex 1

**5. Eligibility and rules of origin**

Participation is open on equal terms to private business entities (companies).

Offeror shall not be considered eligible if: (Ethics Clauses)

1. Any attempt by a candidate or Offeror to obtain confidential information, enter into unlawful agreements with competitors or influence the Commission or the Contracting Authority during the process of examining, clarifying, evaluating and comparing offers will lead to the rejection of his candidacy or offer and may result in administrative penalties.
2. Without the Contracting Authority's prior written authorization, a Contractor and his staff or any other company with which the Contractor is associated or linked may not, even on an ancillary or subcontracting basis, perform other services, carry out works or supply equipment for the project. This prohibition also applies to any other projects that could, owing to the nature of the contract, give rise to a conflict of interest on the part of the Contractor.
3. When putting forward a candidacy or offer, the candidate or Offeror must declare that he is affected by no potential conflict of interest, and that he has no particular link with other Offerors or parties involved in the project. Should such a situation arise during performance of the contract, the Contractor must immediately inform the Contracting Authority.
4. The Contractor must always act honourably and impartially in accordance with the code of conduct of his profession. He must refrain from making public statements about the project or services without the Contracting Authority's prior approval. He may not commit the Contracting Authority in any way without its prior written consent.
5. For the duration of the contract, the Contractor and his staff must respect human rights and undertake not to violate the political, cultural and religious mores of the recipient state.
6. The Contractor may accept no payment connected with the contract other than that provided for therein. The Contractor and his staff must not exercise any activity or receive any advantage inconsistent with their obligations to the Contracting Authority.

7. The Contractor shall refrain from any relationship likely to compromise his independence or that of his staff. If the Contractor ceases to be independent, the Contracting Authority may, regardless of injury, terminate the contract without further notice and without the Contractor having any claim to compensation.
8. The Contracting authority reserves the right to suspend or cancel project financing if corrupt practices of any kind are discovered at any stage of the award process. For the purposes of this provision, "corrupt practices" are the offer of a bribe, gift, gratuity or commission to any person as an inducement or reward for performing or refraining from any act relating to the award of a contract or implementation of a contract already concluded with the Contracting Authority.

## **6. General terms**

- **Delivery and training must be completed at the latest by October 31, 2025.**
- The offer shall be signed by a person, or persons authorized to sign on behalf of the offeror using the Form provided in Annex II.
- The currency of donation is the Euro; therefore, prices of the offers should be expressed in EUR and without VAT.
- The Balkan Medical Task Force – Participating Nation Albania – is the recipient in this Invitation notice for submission of offers. The delivery location is:

**Military Medical Unit  
Zall Herr, 1048 Tirana,  
Republic of Albania**

- The Contracting Authority reserves the right to the final selection based on the funds available for the procurement and other specific requirements.
- Offeror in the offered technical specification must indicate a warranty deadline and potential options for extended warranty.
- Joint Venture, Consortium or Association or any other form of cooperation is not eligible for this contract.
- The Offer and all correspondence and documents related to this procedure exchanged by the Offeror and the Contracting Authority shall be written in English, which is the language of this procedure. Also, all correspondence for payments including pro-forma invoices, and final payment certificates shall be sent to the Contracting Authority in English or local language with English translation.
- This Invitation notice for submission of offers is governed by the laws of the Republic of Slovenia.
- The Offeror shall bear all costs associated with the preparation and submission of his Offer. In no case the Contracting Authority will be responsible or liable for such costs, regardless of the process or results of the Invitation notice for submission of offers procedure.
- The Offeror shall organize and deliver the goods in line with INCOTERMS – DAP (Delivered at Place) to the above stated delivery location and recipient.
- Offeror shall quote all components of the Cost-Breakdown Bill of Quantities exclusive of taxes, customs and import duties. Any item without a price shall not be paid.
- Donor's specific conditions and requirements must be taken into consideration and fulfilled at all related administrative procedures.

- The Contracting Authority reserves the right to accept or reject any offer and/or cancel the whole process of offer selection and reject all offers. The Contracting Authority may do so at any time before informing the selected without any liability to the Offerors. The ITF reserves the right to initiate a new offer selection procedure.

**7. The offer must include all the following information:**

1. Form of Offer (Annex 2)
2. Affidavit Statement (Annex 3)
3. Extract from the register of companies
4. A Certificate of solvency (Bank statement) from the bank in which the transaction account is opened, which shall indicate that the current account was not blocked in the last 6 months on any grounds.
5. Statement on Non-Association (Annex 4)
6. Reference List (supported by relevant official documentation / confirmations) - list all the contracts/projects for the last twenty-four (24) months completed by your company, stating contract value for each, and the month and year completed.

**8. Quality Control and Equipment Handover**

The quality control and equipment handover will be agreed and coordinated between the selected Offeror, ITF, and end recipient Republic of Albania, Participating Nation of Balkan Medical Task Force upon arrival of equipment to the end destination.

**9. Evaluation of Offers**

Acceptable offers will be evaluated by the Contracting Authority applying the criteria of technical suitability, lowest price, and delivery deadline.

**10. POC's**

Please direct technical questions (if any) regarding this Invitation to:

Mr. Iztok Hočević, Email: [Iztok.hocevar@itf.si](mailto:Iztok.hocevar@itf.si)

Please direct administrative/procedural questions (if any) regarding this Invitation to:

Mr Iztok Hočević, Project Manager

E-mail: [Iztok.hocevar@itf.si](mailto:Iztok.hocevar@itf.si)

**11. Offer validity**

The offered prices shall be valid for a period of minimum 90 (ninety) days after the deadline for submission of offers, and must be accompanied by an Offer Guarantee (issued by offerors bank) in the amount of 900 EUR. This guarantee will be released to unsuccessful Offeror(s) once the procedure has been completed and to the successful Offeror upon signature of the contract by all parties.

**12. Advance Payment Guarantees**

The successful contracted party will be asked to provide the advanced payment guarantee, representing 100% of advance payment value, at the signing of the contract. The percentage of the advanced payment bank guarantee will be determined upon the advanced payment amount set in the contract. The advanced payment bank guarantee shall be valid from the date of the signing of the contract and until 90 days after the goods are delivered, installed and training conducted (if applicable). If the selected company fails to provide such bank guarantee, the contract will be void and a new contract may be drawn up and sent to the company which has submitted the second-best admissible offer.

**13. Deadline for submission of offers**

The offers must be received by **ITF Enhancing Human Security, on email E-mail: [Iztok.hocevar@itf.si](mailto:Iztok.hocevar@itf.si)** no later than **Friday, 27 December 2024 by 14:00 hrs (local time)**. Any offer delivered or received after this deadline will not be considered for further evaluation.

**14. Appeal**

After receipt of the notification letter on un-successful offer, the un-successful Offeror can complain to the decision of the Contracting Authority. The Appeal must be submitted in writing and be delivered within the

deadline as stated in the notification letter. The Appeal must be addressed to the ITF Director and must be sent with registered mail or be delivered in-person. The confirmation of payment in amount of 500 EUR or equivalent in local currency must be enclosed to the Appeal. The payment must be made on the ITF account, obtained on [www.itf.si](http://www.itf.si). In the case of justified Appeal, the amount will be returned to the Offeror.

**Annex I:** List of Equipment with detailed technical specifications

**TITLE: Dental Module**

**INSTRUCTIONS:**

In columns No.2 are given **requirements of contracting authority**. Columns No.3 and No.4 are mandatory for Offeror:

- **In column No.3** Offeror must write **MANDATORY** answer on contracting authority requirements from columns No.2. Full description and other technical specifications (**Numbers/Quantity data, description, evidence, delivery schedule, pricing including transportation and training for users – the costs of all activities, certificate, written statement**,);
- **In column No.4** Offeror specifies the number of **attachments or pages**, under which are attached technical specifications and documentation which confirm the requirements of contracting authority, **(MUST BE CLEARLY/APPROPRIATELY MARKED)**.

Offeror must in the offer as proof for listed references in the column No. 2, submit manufacturers publications with technical specification and user data in English and Albanian language for the goods offered (if publications are in any other language translation in English or Albanian language must be submitted) and visual materials (if such visuals exist). In publications and visual material agreement binds must be clearly in accordance with all the demands. Provider must fulfil the demands.

**Offeror's references in the table are binding for the offeror. References/Data/Information from the table will be used and taken in account for the evaluation and the analysis.**

**TITLE: Dental Module**

<b>SUPPLIER</b>	
<b>BRAND OR TYPE</b>	
<b>OFFERED BRAND AND MODEL (Complete designation)</b>	

**Equipment Items and Quantities:**

Dental Module – 1 set

**Technical Specification – Dental Module****1. Table No. 1 - GENERAL DETAILS/CRITERIA**

No.	DESCRIPTION	INFORMATION	ATTACHMENTS
		Obligatory Answer of Supplier – Actual Information	Contract and Technical Specification Attachments (Annex Clarity)
	2	3	4
1.	<b>GENERAL REQUIREMENT:</b>		
1.1.	Dental Module is intended for immediate dental treatment in field conditions. Simple and efficient. It must be independent and highly mobile. No need of fixed infrastructure. Dental Module must be functionally operable.		
1.2.	Portable Dental Unite with three-way syringe, electric micromotor with LED light and 30 programs, electric scaler, with LED, instrument houses detachable, Spray adjustment regulators for each instrument, integrated spray water system for syringe and instruments, Spray water tank, High-performance suction hoses, adapters, suction fluid container, pressure regulator-filter, Sturdy, air and water-tight case, front cover, disc type foot control, integrated silent compressor) with trolley for transport of Dental		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier – Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
2	3	4	
	Unite and rollable or portable case for accessories and equipment. Dental LED light, Dental portable X-Ray with detectors and workstation, Denta vac, Dental cart, Mobile Dental chair, Dental doctor stool, sensors, sensors holders, mobile workstation, portable Autoclave, water distiller, Sealer or sealable pouches and reels, small dental appliances, and tools.		
<b>1.3.</b>	All required dental systems shall be logically organized and ready for field hospital.		

## 2. Table No.2 - STANDARDS AND TECHNICAL REQUIREMENTS

### 2.1. Dental Module

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
2.	REGULATIONS AND RULES IN ALBANIA AND EU		
2.1	In conformance/accordance with regulations which are set for Dental Module in Albania (ALB) and EU.  Supplier delivers quantity and quality certificate- <b>certificate of conformity</b> - homologation		
3.	GENERAL OVERVIEW		
3.1	Dental Unite Dental Module is intended for immediate dental treatment in field conditions. Simple and efficient. It must be independent and highly mobile. No need of fixed infrastructure. Dental Module must be functionally operable with dental equipment.		
4.	OPERATIONAL REQUIREMENTS		
4.1	Conditions for use		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
4.1.1	Dental Module for demanding fieldwork conditions in action anywhere. In conformance/accordance with customer technical conditions and regulations listed in this document. In conformance/accordance with customer technical conditions and regulations listed in this document		
<b>4.2</b>	<b>DIMENSIONS OF DENTAL MODULE</b>		
4.2.1	380mm × 205mm × 530mm, dealer in that country to purchase the appropriate power cord		
4.2.2	Trolley for transport of accessories and dental unit: 410mm × 400mm × 550mm		
4.2.3	Dental accessories case: 385mm × 330mm × 440mm		
4.2.4	Dental LED light. Allowed mounting to the dental chair.		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
4.2.5	Mobile dental chair:  580mm x 380mm x 650mm		
4.2.6	Dental doctor chair: Height adjustable: 540mm–670mm  Diameter Seat: 340mm		
4.2.7	Dental portable X-Ray with detectors and workstation: W 140 mm x L 280 mm x H 270 mm, dealer in that country to purchase the appropriate power cord		
	Sensors Size 1:  39,7 x 25,1 mm		
	Sensors Size 2:  44,1 x 30,4 mm		
	Sensor holders		
4.2.8	Mobile workstation (laptop): 17" Full HD 1920x1080 or better, dealer in that		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	country to purchase the appropriate power cord.		
4.2.9	Portable Autoclave: 600 mm x 550 mm x 550 mm  Dealer in that country to purchase the appropriate power cord.		
4.2.10	Water distiller:  390 x 290 x 290 mm		
4.2.11	Sealer of sealable pouches and reels:  625 x 285 x 145 mm  Dealer in that country to purchase the appropriate power cord		
4.2.12	Dental furniture:  4 m x 0,9 m x 0,6 m		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
4.3	<b>WEIGHT OF DENTAL MODULE</b>		
4.3.1	Weight: Approx. 150 kg		
5.	<b>TECHNICAL REQUIREMENT</b>		
5.1	<b>DENTAL MODULE</b>		
5.1.1	<p><b>Dental unit:</b></p> <p>1 Three-way syringe (air/water/spray) with sterilizable tips (Hose and tips detachable)</p> <p>1 Electric micromotor with LED light and 30 programs (0–40,000), dealer in that country to purchase the appropriate power cord.</p> <p>1 Electric scaler with LED light</p> <p>Instrument hoses detachable</p> <p>Spray adjustment regulators for each instrument</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Integrated spray water system for syringe and instruments (Check valve prevents the back flow of water)</p> <p>1 Spray water tank: 0.75 litres</p> <p>1 High-performance suction hose 300 l/min. Continuously adjustable,</p> <p>Adapter for saliva ejector tips and oral-dry-cup</p> <p>1 Suction fluid container: 1 litre (automatic overflow protection)</p> <p>Pressure regulator-filter</p> <p>Sturdy, air- and water-tight case, will not scratch, dent, or corrode.</p> <p>Detachable front cover</p> <p>Disc type foot control</p> <p>Integrated silent compressor – oil-free: Max. Pressure: 4 bar</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Technical Data</p> <p>Operating pressure: 3–4 bar</p> <p>Each instrument is individually adjustable.</p> <p>Weight up to 13 kg</p> <p>Dimensions: 380mm × 205mm × 530mm</p> <p>Warranty: 2 years</p>		
5.1.2	<p><b>Trolley for transport of accessories and dental unit:</b></p> <p>Features</p> <p>Telescopic handle can be pulled out for transport on wheels or used as a carrying handle to move the trolley.</p> <p>4 Drawers with freely selectable partitioning and magnetic cover</p> <p>1 Storage compartment in the lid</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>5 Wheels (3 castors with brake)</p> <p>Enables also transport of dental unit.</p> <p>Technical Data</p> <p>Weight up to 8 kg</p> <p>Dimensions: 410mm × 400mm × 550mm</p> <p>Warranty: 2 years</p>		
5.1.3	<p><b>Dental accessories case:</b></p> <p>Features</p> <p>1 Handle</p> <p>4 Drawers with freely selectable partitioning and magnetic cover</p> <p>1 Storage compartment in the lid</p> <p>4 Ports for attachment to the trolley</p> <p>Technical Data</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Weight up to 6 kg</p> <p>Dimensions: 385mm x 330mm x 440mm</p> <p>Warranty: 2 years</p>		
5.1.4	<p><b>Dental LED light:</b></p> <p>Light intensity: Max. 70,000 lux at 70cm</p> <p>5 Dimming levels: Operated via sensor or keyboard.</p> <p>CRI: &gt; 95</p> <p>Colour temperature: 4500K (daylight quality)</p> <p>Service life (LEDs): 70,000 hours</p> <p>Housing made of the highest quality materials with gloss-coated, absolutely homogeneous surfaces for perfect hygiene</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Outstanding light intensity and light quality</p> <p>Sensor with five dimming levels</p> <p>Variable light field</p> <p>No infrared radiation, which dries out wounds and slows the healing Process.</p> <p>No UV radiation</p> <p>High reliability and no maintenance costs</p> <p>Unparalleled service life of over 50,000 hours</p> <p>Environmentally friendly (recyclable, does not contain any heavy metals)</p> <p>No risk of implosion (compared to fluorescent tubes)</p> <p>Safe because of the low voltage</p> <p>Protective suitcase for transport</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	Allowed mounting to the dental chair.		
5.1.5	<p><b>Mobile dental chair</b></p> <p>Allows setting up or stow away the chair within 30 seconds.</p> <p>Moving like a trolley: After folding, the chair can be easily pulled with the telescopic handle</p> <p>Comfortable upholstery with premium covering</p> <p>The smooth backrest adjustment is stepless and synchronous with the automatic largest adjustment.</p> <p>Double articulated headrest adjustment</p> <p>Max. patient weight at least 140 kg</p> <p>Weight up to 20 kg</p> <p>Armrest 2pcs</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	Arm with tray  Arm for dental light  Dimensions: 580mm × 380mm × 650mm  Warranty: 2 years		
5.1.6	<b>Dental doctor chair</b>  Chair is set up or stowed away in less than 30 seconds.  To comfortably transport, it can be stowed away in a convenient carrying bag.  Height adjustment by means of a gas spring.  Comfortable upholstery with premium covering  Technical Data  Weight up to 6 kg		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Diameter Seat: 340mm</p> <p>Height adjustable: 540mm–670mm</p>		
5.1.7	<p><b>Dental portable X-Ray with detectors and workstation</b></p> <p>Handheld portable X-ray: Portable and easy to use. Shield must be attached to create safe zone for the operator with internal shield and backscatter shield.</p> <p>Battery 21.6 V DC Lithium Ion</p> <p>Voltage 60 kV true DC</p> <p>Current 2.5 mA</p> <p>Exposure time range 0.02 – 1.00 s</p> <p>Focal spot 0.4 mm</p> <p>Inherent filtration &gt; 1.5 mm Al equivalent</p> <p>Source to skin distance 20 cm</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>X-ray field 60 mm round</p> <p>Maximum duty cycle 1:60</p> <p>Total weight up to 2.7kg</p> <p>Approximate dimensions W 14 cm x L 28 cm x H 27 cm</p> <p><b>Sensors:</b></p> <p>Size 1</p> <p>Sensor size 39,7 x 25,1 mm</p> <p>Active area at least 30,6 x 20,7 mm</p> <p>Resolution at least 2040 x 1380 px</p> <p>Size 2</p> <p>Sensor size 44,1 x 30,4 mm</p> <p>Active area at least 36,0 x 26,1 mm</p> <p>Resolution at least 2400 x 1740 px</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>True resolution more than 20 lp/mm</p> <p>High quality CMOS detector</p> <p>Ethernet connection</p> <p>2 year warranty for detectors</p> <p><b>Sensor holders:</b></p> <p>At least 10 of each (AP, Posterior, Bitewing)</p> <p>Desinfectable</p> <p>Protective bags at least 500 pcs</p> <p><b>Software:</b></p> <p>Image acquisition, Patient management, Image manipulation, Contrast, Brightness, Sharpness, Invert, Noise removal, Crop, Rotate, Mirror</p> <p>Measurements (distances, angles)</p> <p>Annotations</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Marking of the teeth in the database</p> <p>Export formats: DICOM, JPG, PNG</p> <p>TIFF, Importing images.</p> <p>Sending images via email</p> <p>Export with viewer</p> <p>To hard drive, To CD</p> <p><b>Mobile workstation (laptop)</b></p> <p>Processor Intel Core i3 or better</p> <p>RAM 6GB or better</p> <p>Disk SSD 265 GB or better</p> <p>Display/Monitor 17" FullHD 1920x1080 or better</p> <p>Operating system Windows 10 Pro (64 bit) or newer</p> <p>1 year warranty</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	All in protective suitcase with foam inserts.		
5.1.8	<p><b>Portable Autoclave</b></p> <p>Convenient and easy loading with door opening to 110°</p> <p>Assorted water filling and drainage methods (top, front or automatic)</p> <p>Range of automatic one-button programs</p> <p>Remote Wi-Fi monitoring via smartphone or tablet</p> <p>Printer for sterilization process</p> <p>Overall Dimensions up to (D x W x H)</p> <p>60 cm x 50 cm x 55 cm (+/- 10%)</p> <p>Chamber Volume from 22 to 27 Liters</p> <p>Tray Capacity for 5 Trays</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Connections: 1 ethernet port, USB ports, Wi-Fi, integrated automatic water filling and drain connection</p> <p>Item Supplies</p> <p>Drain tube, tray holder.</p> <p>Power Supply &amp; Frequency</p> <p>220V- 230V, 60Hz, dealer in that country to purchase the appropriate power cord.</p> <p>Weight up to 55 kg.</p> <p>All in protective suitcase with foam inserts.</p>		
5.1.9	<p><b>Water distiller</b></p> <p>Stand-alone unit efficiently distils 4 Liters of water in approximately 3.5 hours.</p> <p>Safe access with Hinged condensing lid and condensation contained within a sealed chamber.</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Manufactured with high quality plastics and Teflon heating base for added durability &amp; safety.</p> <p>Automatic shut-off when distillation cycle is completed. It must be easy to clean.</p> <p>Cleaner pack included to ensure boiling chamber is kept clean &amp; free of mineral deposits</p> <p>Filling Volume 4 Liter</p> <p>Distillation Capacity 1 Liter per hour</p> <p>Dimensions - mm</p> <p>(H x W x D) 390 x 290 x 290 (+/- 10%)</p> <p>Weight 4 Kg (+/- 10%)</p> <p>Power Output 500 W</p>		
5.1.10	<p><b>Sealer of sealable pouches and reels</b></p> <p>Rotary sealing device</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Validatable</p> <p>Documentation interface: 2x USB port, 1x Ethernet</p> <p>Device dimensions (WxDxH): 36 x 16 x 17,5 cm (+/- 10%)</p> <p>Weight: 5,6 kg (+/- 10%)</p> <p>Electrical connection: 230 – 240 V / 50/60 Hz</p> <p>Power input: 650 Watts</p> <p>Pre-heating time: 100 sec</p> <p>Sealing duration / speed: 5-15 m/min</p> <p>Temperature range: 100 – 215 degrees Celsius</p> <p>Seal seam width: 12 mm</p> <p>LCD display for setting the device</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	Switch-off tolerance +/- 5°C (DIN 58953-7)  Working table with easy running casters		
5.1.11	<b>Dental furniture:</b>  Lifted from the floor on the feet in stainless steel. The mobile unit trolley with drawers can be easily moved or placed in the dedicated compartment. The drawers have total extraction and are equipped with practical trays with space for tools and materials. With build in sink and facet. Easy to clean and resistant to disinfection.  Dimensions approx.: L 4 m x H 0,9 m x W 0,6 m		
5.1.12	<b>Small dental appliances and tools</b>  Endo motor with apex locator radar (1pcs)		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information/Data	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<p>Surgical loupes (2 pcs)</p> <p>Dental electro bisturi (1 pcs)</p> <p>Piezo ultrasonic scaler (1pcs)</p> <p>Paediatric Forceps kit (1 set)</p> <p>Sets for dental examination (30 sets)</p> <p>Surgical dental kit (2 sets)</p> <p>Dental root elevator (3 sets)</p> <p>Extraction forceps sets (3 sets)</p>		

**3. Tabel No. 3: DOCUMENTATION OF THE DENTAL MODULE, TRAINING, WARRANTY DEMENDS AND in REFERENCES**

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
6.	<b>DOCUMENTATION (by delivery)</b>		
6.1	<p>Supplier and customer shall agree when and how Instructions for use and basic maintenance of the Dental Module will be delivered. Instructions for software and added will be <b>part of the delivery</b> of the Dental Module. Instruction for software and added equipment must be also part of the delivery of the Dental Module.</p> <p>Instructions must be in Albanian language or English language. Content must cover:</p> <ul style="list-style-type: none"> <li>- index,</li> <li>- warnings and hazards,</li> <li>- technical specifications of the Dental Module,</li> </ul>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<ul style="list-style-type: none"> <li>- description of the Dental module and functioning of systems,</li> <li>- Dental Module handling,</li> <li>- description of Dental Module and software and functioning of systems and networks</li> <li>- handling of the medical upgrade,</li> <li>- defects and disorders (tips and help, failure response),</li> <li>- categories and qualities of lubricants and technical liquids,</li> <li>- daily check (check before use, between use and after use),</li> <li>- Dental Module preparation (cleaning, washing, lubrication, adding lubricants and technical liquids) with the purpose to have Dental Module in constant readiness for use,</li> </ul>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<ul style="list-style-type: none"> <li>- conditions of implementation of regular services and checks with detailed definition of necessary contents of services,</li> <li>- List of authorized services for Dental Module and added equipment.</li> </ul> <p>Instructions must be in English language and delivered in printed version (No. of Dental module + 4 copies) and in electronic version (4 copies) in file which is supported in Windows system.</p>		
6.2	<p>Spare parts list of Dental Module and additional equipment and tools. Catalogue must be in English language and delivered in printed version (No. of Dental module + 4 copies) and in electronic version (4 copies) in file which is supported in Windows system.</p> <p>Catalogue must enable availability to web database for spare parts, <b>Dental module information and software with equipment</b></p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
7.	<b>TRAINING (by delivery)</b>		
7.1	<p>Supplier will in coordination with customer by delivery organize training for Dental Module handling and training for basic maintenance of the Dental module, upgrade and equipment.</p> <p>Price of training must be part of purchase price of the Dental Module. Place and date of training shall be agreed between supplier and customer. Training and instruction must be provided in a language that users can understand. Training and instruction must be performed in Albanian language or in language that users can understand. Supplier shall provide certificates which confirm qualified vehicle handling and its equipment</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	<b>Supplier shall provide provisional timeframe of training within his first offer</b>		
8.	<b>WARRANTY AND SUPPORT</b>		
8.1	<p><b>General warranty:</b></p> <p>For DENTAL MODULE <b>min 24 Months</b> with Portable Dental Unite which includes (three-way syringe, electric micromotor with LED light and 30 programs, electric scaler, with LED, instrument houses detachable, Spray adjustment regulators for each instrument, integrated spray water system for syringe and instruments, Spray water tank, High-performance suction hoses, adapters, suction fluid container, pressure regulator-filter, Sturdy, air and water-tight case, front cover, disc type foot control, integrated silent compressor) with trolley for transport of Dental Unite and rollable or portable case for accessories and equipment. Dental LED light, <b>Dental</b> portable X-Ray with detectors and</p>		

No.	DESCRIPTION	INFORMATION Obligatory Answer of Supplier - Actual Information	ATTACHMENTS Contract and Technical Specification Attachments (Annex Clarity)
1	2	3	4
	workstation, Mobile Dental chair, Dental doctor stool, sensors, sensors holders, portable Autoclave, water distiller, Sealer or sealable pouches and reels. For workstation (Laptop) min 12 months from the day of quality check by the delivery of the equipment. Without any restrictions, exceptions, assumptions, or additional conditions. Warranty is valid in EU and other European countries and Albania		

**IMPORTANT:** In points of technical specification where term » in confirmation/accordance » is referred (Dental Modul and its equipment must have manufacturers certificate of conformity which affirms its compliance with the relevant EU standards; Supplier must attach documents/certificates which prove affirms its compliance with the relevant EU standards: Equipment and Consumables) the offeror must write answer YES, if the offer meets the requirements. Relevant certificates must be part of the delivery, and they will be proofed and checked on the delivery day and quality and quantity check. Offeror can attach relevant certificates to the proposal already in the Tender process.



Annex 2

**FORM OF INVITATION FOR SUBMISSION OF OFFERS**

**PROCUREMENT AND DELIVERY OF DENTAL MODULE**

**Date and Place:** \_\_\_\_\_

**1 SUBMITTED by**

<b>Name of Offeror</b>	<b>Nationality</b>

**2 CONTACT PERSON (for this offer)**

<b>Name</b>	
<b>Company</b>	
<b>Address</b>	
<b>Telephone</b>	
<b>Fax</b>	
<b>e-mail</b>	

We, the undersigned, hereby declare that:

1. This Offer is valid for a period of **90 days** from the offer submission deadline;
2. We agree to abide by the ethics Clauses in the Invitation notice for submission of offers and, in particular, have no potential conflict of interests or any relation with other candidates or other parties in this procedure at the time of the submission of this offer;
3. We will inform the Contracting Authority immediately if there is any change in the above circumstances at any stage during the implementation of the contract. We also fully recognize and accept that any inaccurate or incomplete information deliberately provided in this application may result in our exclusion from this and other contracts funded by the Contracting Authority.

Name and first name: [.....]

Duly authorized to sign this Offer on behalf of:

[.....]

Place and date: [.....]

Stamp of the Organization/Legal entity:

This Invitation for submission of offers includes the following annexes:

*[Numbered list of annexes with titles]*

**Annex 3**

**FORM OF AFFIDAVIT STATEMENT**

**Contract No: ITF- \_\_\_\_\_**

**STATEMENT**

With moral and legal responsibilities, I can confirm that company/organization:

\_\_\_\_\_ (Name of the Company/Organization) are not bankrupt or being wound up, are having their affairs administered by the courts, have entered into an arrangement with creditors, have suspended business activities or are in any analogous situation arising from a similar procedure provided for in national legislation or regulations;

\_\_\_\_\_ (Name of the Company/Organization) are not the subject of proceedings for a declaration of bankruptcy, for winding-up, for administration by the courts, for an arrangement with creditors or for any similar procedure provided for in national legislation or regulations;

\_\_\_\_\_ (Name of the Company/Organization) declare that, it is affected by no potential conflict of interest, and that he has no particular link with other Offerors or parties involved in the project. Should such a situation arise during performance of the contract, the Contractor must immediately inform the Contracting Authority.

\_\_\_\_\_ (Name of the Company/Organization) have not been convicted of an offence concerning professional conduct by a judgment which has the force of *res judicata*;

\_\_\_\_\_ (Name of the Company/Organization) are not guilty of grave professional misconduct proven by any means which the Contracting Authority can justify;

\_\_\_\_\_ (Name of the Company/Organization) have fulfilled obligations relating to the payment of social security contributions in accordance with the legal provisions of the country where they are established;

\_\_\_\_\_ (Name of the Company/Organization) have fulfilled obligations relating to the payment of taxes in accordance with the legal provisions of the country where they are established;

\_\_\_\_\_ (Name of the Company/Organization) they are not guilty of serious misrepresentation in supplying the information required by the Contracting Authority as a condition of participation in an invitation to offer or contract;

\_\_\_\_\_ (Name of the Company/Organization) have not been declared to be in serious breach of contract for failure to comply with obligations in connection with another contract with the same Contracting Authority or another contract financed with Community funds;

\_\_\_\_\_ (Name of the Company/Organization) are not in one of the situations allowing exclusion referred to in the Ethics Clauses in connection with the offer or contract.

If it is determined that the above statements are not accurate, we agree that the offer may be declared inadmissible by Contracting Authority, as administrative noncompliant.

Name and Surname: \_\_\_\_\_

Duly authorized to sign this offer:

Signature:

[.....]

Place and date:

Stamp of the Company/Organization:

**Annex 4**

**STATEMENT OF NON-ASSOCIATION**

Offerors should here include a statement that the Offeror (including subcontractors) is not associated, nor has been associated in the past, directly or indirectly, with any other entity that has prepared specifications and documents for the Project or being proposed as Supervisor for the Contract (as stated in Annex 3).

Signature.....

*(a person or persons authorized to sign on behalf of the Offeror)*

Date .....

**Annex 5**

**POWER OF ATTORNEY**

The undersigned, \_\_\_\_\_ (name and surname), being the authorized signatory of the Offeror ..... (name of the company), I am committing that the Offeror will act in accordance with the offer for the submitted Offer Reference No. : **ITF -** , for the contract No. .....

Signature:

Position:

Stamp of the Offeror:

Date: